



PHILIPPINE HEART CENTER
East Avenue, Quezon City

SUPPLEMENTAL BID BULLETIN No. 062-2019

SUBJECT : Modification in the Bidding Document

PROJECTS : 3 units Transport Cardiac Monitor - ITB No. 064.19

To ensure that the transactions are comparatively advantageous to the interests of the PHC, the originally scheduled Project as mentioned above, modifications are hereby introduced:

Item No.	As Published, PhilGeps (From)	Corrected (should be read as) (To)
1	With stand parameter we lead ECG, NIBP, SPO2, RR, temp, with provision for a pressure line dedicated cables and connection	With stand parameter at least 6 lead ECG, NIBP, SPO2, RR, temp, with provision for a pressure line dedicated cables and connection
2	Configurable alarm settings and alarm history	Configurable alarm settings and alarm history audible-visual
7	e. Light weight maximum of 5 lbs	8. Monitor: Light weight maximum of 3-5 kgs
8	With a minimum of warranty of 5 years on parts and services	9. With a minimum of warranty of 5 years on parts and services
9	Ten (10)years guarantee on availability of parts	10. Ten (10)years guarantee on availability of parts

This Supplemental Bid Bulletin is hereby issued to modify the original parameters/advertisement as stated therein.

Please be guided accordingly.

Approved. Done this 13th day of August 2019, BAC Office, PHC.


MARIETTA A. VELASCO, RN, MAN
Chairman, BAC for Pharma Supplies and Med Equipment